

## **Declaration of Conformity**

Date: 17 December 2018

Manufacturer/Place of Declaration:

Address:

Nonin Medical, Inc.

13700 1st Avenue North

Plymouth, MN 55441-5443 USA

Model Numbers and

3150USB, 3150 SK USB (CFG)

**Product Designations:** 

WristOx2® Wrist-Worn Pulse Oximeter with Bluetooth®

Device Category(ies):

Pulse oximeter, battery powered

GMDN Number(s):

45607

Date Added:

June 2017

We herewith declare that the above mentioned pulse oximeter is classified as Class IIb (using rule 10) and complies with Annex II, section 3 of the Directive No. 93/42/EEC on medical devices, as amended (2007/47/EC), and all applicable clauses of ISO 80601-2-61, Medical electrical equipment -- Particular requirements for the basic safety and essential performance of pulse oximeter equipment for medical use.

This declaration applies to all devices manufactured from the date of issuance until it is either superseded by another declaration or withdrawn.

We further declare, under our sole responsibility, that the above mentioned pulse oximeter to which this declaration relates is in conformity with the following standards and/or other normative documents:

DIRECTIVE 2014/53/EU OF THE EUROPEAN PARLIAMENT AND OF THE COUNCIL of 16 April 2014 on the harmonisation of the laws of the Member States relating to the making available on the market of radio equipment and repealing Directive 1999/5/EC

EN 301 489-17 V3.1.1:2017, ElectroMagnetic Compatibility (EMC) standard for radio equipment and services; Part 17: Specific conditions for Broadband Data Transmission Systems; Harmonised Standard covering the essential requirements of article 3.1(b) of Directive 2014/53/EU<sup>1</sup>

Notified Body:

TÜV SÜD Product Service GmbH

Ridlerstrasse 65 D-80339 München

Germany

**EC Certificate Number:** 

G1 024497 0030 Rev.00

Number

CE0123

Signature:

Name:

Title:

Principal Regulatory Affairs Specialist

Authorized EC Representative:

Medical Product Service (MPS) GmbH

Borngasse 20 35619 Braunfels

Germany

<sup>1</sup> Element NONN0179, NONN0179.1